

PORTABLE OXYGEN CONCENTRATORS

DOCUMENT REVISED NOVEMBER 2016

FAA APPROVED AS OF AUGUST 23, 2016



AEROMEDIC.COM
1-800-346-3556

PICTURES ARE NOT TO SCALE. CHECK YOUR AIRLINES WEBSITE FOR RULES ABOUT FLYING WITH PORTABLE OXYGEN CONCENTRATORS.

CONSULT YOUR DOCTOR ABOUT FLYING WITH OXYGEN. NEED MORE CARDS LIKE THESE? EMAIL INFO@AEROMEDIC.COM



Invacare XPO2M™
Weight: 6 lbs
Dims: 10" x 7" x 4"



AirSep FreeStyle™
Weight: 4.4 lbs
Dims: 8.6" x 6.1" x 3.6"



AirSep FreeStyle 5™
Weight: 6.2 lbs
Dims: 10.7" x 6.6" x 4.4"



AirSep Focus™
Weight: 1.75 lbs
Dims: 6.2" x 4.6" x 2.4"



Precision Medical EasyPulse™
Weight: 6.8 lbs
Dims: 10.1" x 6.5" x 4.5"

AirSep LifeStyle™
Weight: 9.75 lbs
Dims: 5.5" x 7.25" x 16.31"



SeQual Eclipse®
Weight: 17.9 lbs
Dims: 19.3" x 12.3" x 7.1"



Inogen G3™
Weight: 4.8 lbs
Dims: 8.75" x 3.0" x 7.25"



Inogen G2™
Weight: 7 lbs
Dims: 11.61" x 5.98" x 10.75"



Inogen One™
Weight: 9.7 lbs
Dims: 11.62" x 6" x 12.39"



Oxus™
Weight: 9.8 lbs
Dims: 11.6" x 7.4" x 4.6"



DeVilbiss Healthcare iGo™
Weight: 19 lbs
Dims: 15" x 11" x 8"



Inova Labs ActivOx™
Weight: 4.83 lbs
Dims: 9.05" x 7.875" x 4.38"



Inova Labs LifeChoice™
Weight: 4.9 lbs
Dims: 9.5" x 7.5" x 3.125"



VBOX Trooper™
Weight: 4.3 lbs
Dims: 6" x 2.8" x 7.25"



Philips Respironics EverGo™
Weight: 8.5 lbs
Dims: 12" x 6" x 8.5"



Precision Medical™
EasyPulse TOC
Weight: 11.5 lbs
Dims: 14.63" x 10.5" x 7"



SeQual Equinox (US)™
& Oxywell (Japan)™
Weight: 14 lbs
Dims: 13.6" x 10.6" x 7.4"



OxLife Independence™
Weight: 16.7 lbs
Dims: 20.29" x 10.85" x 9.45"



Invacare Solo2™
Weight: 19.9 lbs
Dims: 16.5" x 11" x 8"



Philips Respironics SimplyGo™
Weight: 10 lbs
Dims: 11.5" x 10" x 6"



Philips Respironics™
SimplyGo Mini
Weight: 5 lbs.
Dims: 10.2" x 8.3" x 2.6"



GCE Healthcare Zen-O™
Weight: 10.3 lbs.
Dims: 8.3" x 6.6" x 12.3"



O2 Concepts™
Oxlife Freedom
Weight: 5.5 lbs
Dims: Unavailable at Press Time




Invacare™
Platinum Mobile
Weight: 4.98 lbs
Dims: 9.45" x 7.5" x 3.88"



SeQual SAROS®
Weight: 12.25 lbs
Dims: 26.8" x 4.375"



Inogen G4™
Weight: 2.8 lbs.
Dims: 7.2" x 5.91" x 2.68"

 The manufacturer of this POC has determined this device conforms to all applicable FAA requirements for POC carriage and use on board aircraft.

NEW POC'S WITH APPROVED LABEL

EXAMPLE FAA APPROVED LABEL



PORTABLE OXYGEN CONCENTRATORS

AS OF LATEST FAA RULING FEBRUARY 18, 2014 THIS DOCUMENT REVISED NOVEMBER 2016 DOCTOR RX REQUIRED

FAA APPROVED BATTERY DURATION

	Setting					
	1	2	3	4	5	6
AirSep Focus Focus w/ Battery AirBelt	--	1.5 hrs 4.0 hrs	--	--	--	--
AirSep FreeStyle FreeStyle Battery AirBelt	6.5 hrs 6.5 hrs	3.5 hrs 3.5 hrs	3.0 hrs 3.0 hrs	--	--	--
AirSep FreeStyle 5 FreeStyle 5 Battery AirBelt	4.25 hrs 4.0 hrs	2.5 hrs 2.5 hrs	2.0 hrs 2.0 hrs	1.75 hrs 1.75 hrs	1.25 hrs 1.5 hrs	--
AirSep LifeStyle	50 min	50 min	50 min	50 min	50 min	--
Inogen One	3.0 hrs	3.0 hrs	2.5 hrs	2.5 hrs	2.0 hrs	--
Inogen One G2 12 cell G2 24 cell Extended Battery	4.0 hrs 8.0 hrs	3.5 hrs 7.0 hrs	2.7 hrs 5.4 hrs	2.2 hrs 4.4 hrs	2.0 hrs 4.0 hrs	--
Inogen One G3 8 cell G3 16 cell Extended Battery	6.0 hrs 12.0 hrs	4.5 hrs 9.0 hrs	3.0 hrs 6.0 hrs	2.0 hrs 4.0 hrs	--	--
Inogen One G4 Single Battery G4 Double Battery	3.0 hrs 6.0 hrs	2.0 hrs 4.5 hrs	1.25 hrs 2.5 hrs	--	--	--
Inova Labs LifeChoice Lifechoice External Battery +add	2.0 hrs 3.0 hrs	2.0 hrs 3.0 hrs	2.0 hrs 3.0 hrs	--	--	--
Inova Labs LifeChoice Activox Activox External Battery +add	12.0 hrs 3.0 hrs	6.0 hrs 3.0 hrs	5.0 hrs 2.0 hrs	--	--	--
Inova Labs Activeox Pro Activeox External Battery +add	12.5 hrs 3.0 hrs	7.25 hrs 3.0 hrs	6.25 hrs 3.0 hrs	--	--	--
Inova Labs Activeox 4L Activeox External Battery +add	10.25 hrs 3.0 hrs	8.25 hrs 3.0 hrs	5.0 hrs 3.0 hrs	4.0 hrs 3.0 hrs	--	--
Invacare Platinum Mobile	5.08 hrs	3.5 hrs	2.3 hrs	1.75 hrs	--	--
Invacare Platinum Mobile with 2nd Battery	10.17 hrs	7.0 hrs	4.67 hrs	3.5 hrs	--	--
Invacare XPO2 XPO2 External Battery +add	3.67 hrs 3.5 hrs	2.67 hrs 2.4 hrs	2.0 hrs 2.0 hrs	1.67 hrs 1.5 hrs	1.3 hrs 1.0 hrs	--
O2 Concepts Oxlife Freedom	DATA UNAVAILABLE AT PRESS TIME					
Oxus RS-00400, Delphi RS-00400	3.45 hrs	4.0 hrs	2.5 hrs	2.0 hrs	1.75 hrs	--
Precision Medical EasyPulse EasyPulse External Battery +add	5.5 hrs 5.5 hrs	3.8 hrs 3.8 hrs	2.8 hrs 2.8 hrs	--	--	--
Philips Respironics EverGo EverGo with 2nd Battery	6.0 hrs 12.0 hrs	4.0 hrs 8.0 hrs	3.0 hrs 6.0 hrs	2.0 hrs 4.0 hrs	2.0 hrs 4.0 hrs	2.0 hrs 4.0 hrs
Philips Respironics SimplyGo Mini Philips Respironics SimplyGo Mini Ext. Battery	6.0 hrs 12.0 hrs	4.5 hrs 9.0 hrs	3.0 hrs 6.0 hrs	2.25 hrs 4.5 hrs	2.0 hrs 4.0 hrs	--
DeVilbiss HealthCare iGo Continuous Flow	5.4 hrs 4.0 hrs	4.7 hrs 2.4 hrs	4.0 hrs 1.6 hrs	3.5 hrs --	3.2 hrs --	3.0 hrs --
Precision Medical EasyPulse TOC Continuous Flow	DATA UNAVAILABLE AT PRESS TIME					
GCE Healthcare Zen-O Zen-O with 2nd Battery 0.5 = 3.0/6.0 hrs Continuous Flow Battery / w/ 2nd Battery	4.0 hrs 8.0 hrs 1.75/3.5 hrs	4.0 hrs 8.0 hrs .75/1.5 hrs	3.0 hrs 6.0 hrs --	2.25 hrs 4.5 hrs --	2.0 hrs 4.0 hrs --	1.75 hrs 3.5 hrs --
Invacare Solo2* Continuous Flow 0.5 = 4.5 hrs	4.5 hrs 3.5 hrs	3.5 hrs 2.5 hrs	3.0 hrs 1.5 hrs	2.5 hrs --	2.5 hrs --	--
Oxlife Independence* With 2nd Battery 0.5 = 2.4 hrs Continuous Flow Battery / w/ 2nd Battery	2.8 hrs 5.6 hrs 2.4/4.8 hrs	2.8 hrs 5.6 hrs 1.4/2.8 hrs	1.4hrs 2.8 hrs 53min/1.8hr	1.4 hrs 2.8 hrs --	53 min 1.8 hrs --	--
Philips Respironics SimplyGo* Continuous Flow 0.5 = 2.9 hrs	3.4 hrs 1.6 hrs	3.0 hrs 0.9 hrs	2.2 hrs --	1.7 hrs --	1.3 hrs --	1.3 hrs --
SeQual Eclipse* Continuous Flow 0.5 = 4.4 hrs	4.4 hrs 3.7 hrs	3.6 hrs 2.0 hrs	3.0 hrs 1.3 hrs	2.6 hrs --	2.3 hrs --	2.1 hrs --
SeQual SAROS Continuous Flow	1.2 hrs 1.1 hrs	1.1 hrs 45 min	53 min 30 min	47 min --	46 min --	37 min --
VBOX Trooper	3.5 hrs	3.5 hrs	3.5 hrs	3.5 hrs	3.5 hrs	3.5 hrs

SeQual Equinox (US) & Oxywell (Japan Model)

Pulse Delivery	12 Cell	24 Cell	Continuous Flow	12 Cell	24 Cell
Setting 1.0 Bolus 16 mL	2.827 hrs	6.03 hrs	Setting 0.5 LPM	2.48 hrs	5.28 hrs
Setting 2.0 Bolus 32 mL	2.783 hrs	5.94 hrs	Setting 1.0 LPM	2.16 hrs	4.84 hrs
Setting 3.0 Bolus 48 mL	2.38 hrs	5.39 hrs	Setting 2.0 LPM	1.21 hrs	2.75 hrs
Setting 4.0 Bolus 64 mL	2.31 hrs	4.94 hrs	Setting 3.0 LPM	0.8 hrs	1.84 hrs
Setting 5.0 Bolus 80 mL	2.22 hrs	4.75 hrs			
Setting 6.0 Bolus 96 mL	1.70 hrs	3.85 hrs			
Setting 7.0 Bolus 128 mL	1.56 hrs	3.33 hrs			
Setting 8.0 Bolus 160 mL	1.28 hrs	2.72 hrs			
Setting 9.0 Bolus 192 mL	1.0 hrs	2.15 hrs			

SeQual Eclipse 5

Pulse Delivery	Continuous Flow
Setting .5 NA	Setting .5 4.4 hrs
Setting 1.0 5.4 hrs	Setting 1.0 3.7 hrs
Setting 2.0 5.2 hrs	Setting 2.0 2.0 hrs
Setting 3.0 4.9 hrs	Setting 3.0 1.3 hrs
Setting 4.0 4.0 hrs	Pulse 128 mL 2.5 hrs
Setting 5.0 3.7 hrs	Pulse 160 mL 2.0 hrs
Setting 6.0 3.5 hrs	Pulse 192 mL 1.7 hrs



800-346-3556
AEROMEDIC.COM

All times based on NEW batteries and are estimates as per each manufacturer. Each Airline has its own rules about flying with portable oxygen concentrators.

RENTALS & SALES

24HR SUPPORT